



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001108 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TS4
12-22-2011

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 11/10/11 Cal 1 N/A AC Cal 2 N/A AC

Between the dates of N/A and N/A the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 11/10/11.
The instrument was returned to service on _____.

Sarah Cione

Sarah L. Cione


12/20/11

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 20, day of December, 20 11.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

1 of 2sc

TS4
12-22-2011

Date: 11/10/11

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-001108

Reason For Service: Voltage/Current Fails during Diagnostics

Date Out of Service: 11/10/11

Calibration Check completed prior to removal from service

Date: N/A

SQAP completed prior to removal from service

Date: N/A

Corrective Action:

☒ Sent to CMI, Inc.

Data downloaded prior to sending to CMI

2 of 2sc



Service Evaluation Form

This form **MUST** be completed and enclosed with instrument to be serviced.
Failure to complete and return this form may cause delays in service.

(Note: please ship items in their original shipping container(s) or a similar protective box.)

1. Contact information: Customer Number 105075 (contact Customer Service)
Name Terry Gallegos Phone: (520) 837-7677
Fax: (520) 791-4642 Email: Terry.Gallegos@tucsonaz.gov

2. Bill to Address: Ship to Address:
Finance / Accounts Payable Tucson Police Department Crime Lab
Po Box 27450 1306 W. Miracle Mile
Tucson, AZ 85726-7450 Tucson, AZ 85705

3. Serial Number: 80-001108 Instrument Model: I8000

4. Detailed Description of Problem:
Voltage/Current Fails during Diagnostics

Hazardous Material Warning! - DO NOT return gas cylinders with instrument!

☐ I Authorize Repairs Up To: ☐ All ☐ \$250 ☐ \$500 ☐ \$750 ☐ Other \$ _____
Purchase Order Number (attach a copy of P.O. if applicable) 34770

Authorized By:

Terry Gallegos Toxicology Coordinator
Name (Please Print) Title

[Signature] 11.22.2011
Signature Date

Ship to:

CMI, Inc.
Attn: Service Dept.
316 East Ninth Street
Owensboro, KY 42303

☒ No, please send estimate before repairs are made.

Note: An estimate will be faxed before performing any repairs and may cause delays in service.

An evaluation fee (\$79.00 or actual costs) will apply to estimates that are not repaired.